

E003: Refrigerators and freezers

PQS code: E003/129

Type of appliance: Combined icelined refrigerator/waterpacks freezer

Manufacturer's reference: TCD-100

Manufactured in: China; People's Republic of

Company:Qingdao Aucma Global Medical Co.,Ltd.Address:315 Qianwangang Road, Economic

&Technological Development Zone, Qingdao,

China

Telephone: +86 532 86763602

Email: lishouda @aucma.com.cn

Web address: http://aucmaglobal.com.cn

Specifications

Climate zone: Hot Min rated ambient temp: +5°C

Refrigerant: #R290 Energy source:

Appliance tested at: +43°C Ext dimensions (HxLxD) W1050*D750*H975 mm

Performance at: +43°C Fuel and cycle type: Electric - compression

SOLAR DIRECT DRIVE REFRIGERATOR AND / OR FREEZER Freeze protection: Grade A

Vaccine storage capacity: 48L Gross volume (Liters): 38L

(Liters)

WHO/PQS protocols

Gross volume (Liters): 59L

Waterpack freezing capacity: 2kg/24h

Waterpack storage capacity:

10.7 kg

Autonomy as per 180 h At a solar radiation reference 3.5 kWh/m2/day

period of:

Warning! For solar direct drive units, the correct sizing of the solar panel array for a specific site is complex

and critical. It must be agreed with both the appliance manufacturer and with the Qualified Supplier

of the solar energy system at the time of ordering.

Comments:	"1、48-liter vaccine storage chamber volume; 38-liter water pack freezer chamber volume; 2、Integrated USB temperature data module with 10 year data storage capacity; 3、180 hours of holdover at constant ambient temperature of 43 °C; 4、2.0kg daily Water-pack freezing capacity, 10.7kg water-pack storage compartment capacity; 5、180 hours autonomy at constant ambient temperature of 43 °C; 6、Integrated maximum/minimum temperature and alarm data with 30 day history; 7、Integrated USB charging port 5V / 0.5A;		
Accessories:	"1、2 Vaccine boxes 2、2 Keys 3、1 Funnel 4、Documentation 5、2 baskets		
Spare parts (ref):	000001485: Compressor 3000002307: DC fan 3001000717: Electronic control unit		
Shipping volume:	1.62 m3	Shipping weight:	260 kg
Price year	2022		
Quality standard:	- ISO 9001: 2015 -	Specification ref.	
Verification report:	WB-21-0332	Verification laboratory:	CHEARI
Current PQS status:	pre-qualified: 14 Sep 2022	Validity u	ntil: <i>May</i> 2025

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.